



USP <88> Class Plastic Designation

Sponsor: Garlock Sealing Technologies
1666 Division Street
Newark, NY 14522-9350

Attn: Matt Tones

Test Nos.: T09-1343, 1344, 1345

Test Material: GYLON 3500

This test material meets the requirements for USP <88> Class VI designation at a temperature extraction of 121°C for 1 hour.

Jonia Schaubrock

Study Director

6-4-09

Date

S:\TOX\REPORT\2009\T09-1343, 1344, 1345 Class letter.doc