



USP <88> Class Plastic Designation

Sponsor: Garlock Sealing Technologies
1666 Division Street
Palmyra, NY 14522-9350

Attn: Matt Tones

Test No(s): T10-0691, 0692, 0693

Test Material: One (1) GYLON 3510

This test material meets the requirements for USP <88> Class VI designation at a temperature extraction of 121°C for 1 hour.

Jonia Schaubroek

Study Director

3-23-10

Date